



TOP VIEW



FRONT VIEW



EXTENDED UNIBODY LITTROW LASER -- UPPER LID

E. COOK, P. MARTIN, D. STECK

3/24/11

NOTES:

- A. 8-32 clearance through part ($\text{\O} 0.177''$), for mounting lids to main cavity.
- B. 8-32 tapped through part.
- C. 0.115" wide flat-bottom channel, 0.080" deep, oriented as shown. Channel is for a 3/32" thick, 7" inner diameter O-ring.

ADDITIONAL NOTES:

1. Material is 6061-T6 aluminum. Starting block is 4.48" x 10" x 0.75".
2. Standard tolerance is 0.005" unless otherwise noted.